

# Portable Indoor Air Quality Instruments and Smart Sensor Specifications



Critical Environment Technologies Canada Inc. Unit 145, 7391 Vantage Way Delta, BC V4G 1M3 Canada

www.critical-environment.com

TF: +1.877.940.8741

T: +1.604.940.8741

F: +1.604.940.8745

© August 2018 | Data subject to change without notice.

# **Table of Contents**

Product Overview: Indoor Air Quality Instruments	4
Overview: Gas Detection Sensor Technologies	5
Target Gas Sensor Specifications	
Relative Humidity (RH)	7
Temperature	7
Ammonia (NH <sub>3</sub> )	8
Arsine (AsH <sub>x</sub> )	8
Carbon Dioxide (CO.)	9
Carbon Monoxide (CO)	12
Chlorine (Cl <sub>3</sub> )	13
Chlorine Dioxide (ClO <sub>2</sub> )	14
Combustibles	15
Ethylene (C,H,)	16
Ethylene Oxide (C,H <sub>4</sub> O)	16
Fluorine (F <sub>2</sub> )	17
Formaldehyde (CH <sub>3</sub> 0)	18
Hydrogen(H <sub>2</sub> )	19
Hydrogen Chloride (HCI)	
Hydrogen Cyanide (HCN)	20
Hydrogen Fluoride (HF)	21
Hydrogen Sulphide (H,S)	22
Nitrogen Dioxide (NO <sub>3</sub> )	22
Nitric Oxide (NO)	23
Oxygen (0,)	24
Ozone (0 <sub>3</sub> )	25
Phosphine (PH <sub>3</sub> )	26
Silane (SiH <sub>4</sub> )	27
Sulphur Dioxide (SO <sub>2</sub> )	28
Total Volatile Organic Compounds (TVOC)	29
Particulate Sensor (PM2.5 or PM10)	29
Additional Chemical Symbols	30
Important Notes	30

# **Indoor Air Quality Instruments**

# **YESAIR Eight Channel Air Quality Monitor**



- » 8 sensor capacity
- » > 30 plug & play gas sensor choices
- » PM2.5 or PM10 particulate sensor option
- » Data logging to SD flash card
- » Lightweight, contoured & comfortable handheld device



# **YES Plus LGA Fifteen Channel Air Quality Monitor**



- » 15 sensor capacity
- » > 30 plug & play sensor choices
- » Data logging to SD flash card
- » Internal sample pump with inline filter



# Gas Detection Sensor Technologies

Categories of gas detection systems are defined by the technology they use: electrochemical sensors generally detect toxic gases, catalytic and infrared sensors detect combustible gases and TVOC sensors detect toxic gas, organic compounds and chemicals.



#### **ELECTROCHEMICAL**

Toxic gas sensors & oxygen sensors

#### **CATALYTIC**

Combustible gas sensors, toxic gas sensors at very high concentrations (% volume)

#### LIFE SPAN

2 - 5 years (sensor type & manufacturer dependent). Oxygen typically 3 years.

#### LIFE SPAN

3 - 8 years (typically if not poisoned)

#### SPECIFICITY

Specific to target gas with known cross sensitivity to a small variety of gases.

#### SPECIFICITY

Specific to combustible gases only in the LEL ranges

#### RANGE

Typically 0 - 1.0 ppm or 0 - 2,000 ppm, sensor dependent.

#### RANGE

0 - 100% LEL of target gas

#### POISONING / DAMAGING

- Oxygen depravation
- Exposure to high concentrations of solvent vapours
- Very high concentrations of target gas
- Reactive gases
- Environments with high temperatures, low temperatures (freezing)
- Very low levels of relative humidity (less than 10 - 15%)

#### POISONING / DAMAGING

High concentrations of target gas, lead vapours, silicon vapours, alkylated heavy metals.

#### APPLYING SPAN GAS

- Use span gas with air balance or nitrogen balance
- Flow rate should be a minimum of 0.5 LPM (lighter-than air gases) to a maximum of 1.0 LPM (heavier-than-air gases)
- Do NOT humidify span gas when flowing

#### APPLYING SPAN GAS

- Use span gas with air balance ONLY.
- Flow rate should be a minimum of 0.5 LPM (lighter-than air gases) to a maximum of 1.0 LPM (heavier-than-air gases)
- Do not humidify span gas when flowing

# **Gas Detection Sensor Technologies**

#### **INFRARED**

#### PID / TVOCS

Toxic, combustible & refrigerant gas sensors

Toxic gas sensors, organic compounds & chemicals

#### LIFE SPAN

10 years +

#### LIFE SPAN

3 - 8 years (typically if not contaminated and with regular maintenance)

#### SPECIFICITY

Specific to target gas.

#### **SPECIFICITY**

Non specific. Will respond to any compound that has an ionization potential less than the ionization potential of the lamp.

#### RANGE

0 - 1,000 ppm or 0 - 100% volume. Target gas, manufacturer dependent.

#### RANGE

0 - 30 ppm or 0 - 300 ppm, sensor dependent

#### POISONING / DAMAGING

No known poisoning agents. Condensing humidity will damage sensor and distort readings.

#### CONTAMINATION

Many other gases, vapours, chemicals. Condensing humidity can cause false positive readings.

#### APPLYING SPAN GAS

- Use span gas with air balance or nitrogen balance.
   Nitrogen balance ONLY for CO<sub>3</sub> sensors.
- Flow rate should be approximately 0.5 LPM. Some sensors are flow sensitive.
- Do NOT humidify span gas when flowing.

#### **APPLYING SPAN GAS**

- Use span gas with air or nitrogen balance.
- Flow rate should be a minimum of 0.5 LPM.
- Do NOT humidify span gas when flowing.

# **Target Gas Sensors**

JENJUN	
Туре	Thin film capacitive
Standard Range	5 - 95% RH (non-condensing)
Resolution	2% RH
Accuracy	No data available
Long Term Drift	2% (±) / 12 months
Response Time	< 10 seconds
INSTRUMENT	
Displayed Resolution	1%
Warm Up Time @ Switch On	5 minute operational, 20 minute max accuracy
Recommended Calibration Frequency	1 yr for best performance
ENVIRONMENTAL	
Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	5 - 95% non- condensing
Operating Life (Estimated)	3 yrs +
Temperature	
SENSOR	
Туре	Negative Coefficient Thermistor
Standard Range	0°C to 50°C (32°F to 122°F)
Resolution	0.1°C @ 25°C
Accuracy	No data available
Long Term Drift	0.5°C (±) / 12 months
Response Time	< 10 seconds
INSTRUMENT	
Displayed Resolution	0.1°C
I	

**Releative Humidity (RH)** 

Warm Up Time @ Switch On

**ENVIRONMENTAL Operating Temperature** 

**Operating Humidity** 

Operating Life (Estimated)

**Recommended Calibration Frequency** 

SENSOR

6 yrs +

1 yr for best performance

0°C to 50°C (32°F to 122°F)

5 - 95% non-condensing

5 minute operational, 10 minute max accuracy

Ammonia (NH <sub>3</sub> )	50 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 50 ppm
Resolution	1 ppm
Accuracy	No data available
Long Term Drift	< 5% / 6 months
Response Time	$t_{90} = < 60$ sec calculated fr 5 minute exposure
	Alcohols @ 1,000 ppm = 0 ppm
	$CO_{2} @ 5,000 \text{ ppm} = 0 \text{ ppm}$
	$CO^{2}$ @ 100 ppm = 0 ppm
Cross Sensitivities	Hydrocarbons @ % range = 0 ppm
Closs Sensitivities	$H_{2} @ 10,000 \text{ ppm} = 0 \text{ ppm}$
	$H_{3}^{-}S @ 20 \text{ ppm} = 2 \text{ ppm}$
	Cross sensitivity list not fully completed.
	Sensor maybe sensitive to other gases.
INSTRUMENT	
Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approx 2 - 2.5 min (warm up delay time)
Recommended Calibration Frequency	6 months
ENVIRONMENTAL	
Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	15 - 90% non-condensing

Arsine (AsH <sub>3</sub> )	1 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 1 ppm
Resolution	< 15 ppb @ 20°C (68°F)
Accuracy	No data available
Long Term Drift	< 5% / 6 months
Response Time	$t_{q_0} = < 30$ seconds calc fr 2 minute exposure
	CO @ 85 ppm = 0 ppm
	$H_{2} @ 3,100 \text{ ppm} = 0 \text{ ppm}$
Cross Sensitivities	$NO_{2}$ @ 10 ppm = 2 ppm
	$C_1H_5OH @ 25,000 \text{ ppm} = 0 \text{ ppm}$
	$H_2S @ 18 \text{ ppm} = 10.8 \text{ ppm}$

2 yrs

Operating Life (Estimated)

	SO <sub>2</sub> @ 18 ppm = 5.4 ppm	
	$Cl_{3}@0.85 \text{ ppm} = 0.24 \text{ ppm}$	
	HCl @ 7.8 ppm = 1 ppm	
	HF @ 7.2 ppm = 0 ppm	
Cross Sensitivities continued	HCN @ 12.6 ppm = 0.7 ppm	
	$SiH_{A}$ @ 4.3 ppm = 0.7 ppm	
	H <sub>2</sub> Se @0.8 ppm = 0.24 ppm	
	$B_{2}H_{4} @ 0.2 \text{ ppm} = 0.28 \text{ ppm}$	
	PH <sub>2</sub> @ 0.2 ppm = 0.24 ppm	

Displayed Resolution	0.001 ppm (1 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

# **ENVIRONMENTAL**

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	20 - 95% non- condensing
Operating Life (Estimated)	1.5 yrs

Carbon Dioxide (CO <sub>2</sub> )	5,000 ppm
SENSOR	
Туре	Infrared
Standard Range	0 - 5,000 ppm
Resolution	50 ppm fr 0 - 2,500 ppm, then 100 ppm up to FSD
Accuracy	$\pm$ 2% full scale @ 20°C (68°F), 1 bar pressure, applied gas 2.5% volume CO $_{\odot}$
Long Term Drift	$\pm$ 50 ppm / month @ 20°C (68°F) ambient, (max $\pm$ 150 ppm / yr)
Response Time	t <sub>90</sub> = > 30 seconds @ 20°C (68°F)
Cross Sensitivities	None

# INSTRUMENT

Displayed Resolution	1 ppm
Warm Up Time @ Switch On	Approx 2 - 2.5 min (warm up delay time)
Recommended Calibration Frequency	2 yrs, 1 yr for best accuracy

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	5 - 95% non-condensing
Operating Life (Estimated)	5 - 10 yrs

Carbon Dioxide (CO <sub>2</sub> )	10,000 ppm
SENSOR	
Туре	Infrared
Standard Range	0 - 10,000 ppm
Resolution	50 ppm fr 0 - 2,500 ppm, then 100 ppm up to FSD
Accuracy	$\pm$ 2% full scale @ 20°C (68°F), 1 bar pressure, applied gas
Accuracy	2.5% volume CO <sub>2</sub>
Long Term Drift	± 500 ppm / month @ 20°C (68°F) ambient
Response Time	$t_{q_0} = > 30 \text{ seconds } @ 20^{\circ}\text{C } (68^{\circ}\text{F})$
Cross Sensitivities	None
INSTRUMENT	
Displayed Resolution	1 ppm
Warm Up Time @ Switch On	Approx 2 - 2.5 min (warm up delay time)
Recommended Calibration Frequency	2 yrs, 1 yr for best accuracy
ENVIRONMENTAL	
Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	5 - 95% non-condensing
	-
Carbon Dioxide (CO <sub>2</sub> )	5% volume
SENSOR	
Туре	Infrared
Standard Range	0 - 5% volume
	1% of measuring range for readings above 50% of range

Carbon Dioxide (CO <sub>2</sub> )	5% volume
SENSOR	
Туре	Infrared
Standard Range	0 - 5% volume
Resolution	1% of measuring range for readings above 50% of range,
	0.5% of measuring range for readings below 50% of range
Accuracy	$\pm$ 2% full scale @ 20°C (68°F), 1 bar pressure, applied gas
	2.5% volume CO <sub>3</sub>
Long Term Drift	± 500 ppm / month @ 20°C (68°F) ambient
Response Time	$t_{on} = > 30 \text{ seconds } @ 20^{\circ}\text{C } (68^{\circ}\text{F})$
Cross Sensitivities	None

Displayed Resolution	0.1% volume
Warm Up Time @ Switch On	Approx 2 - 2.5 min (warm up delay time)
Recommended Calibration Frequency	2 yrs, 1 yr for best accuracy

# **ENVIRONMENTAL**

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	5 - 95% non-condensing
Operating Life (Estimated)	5 - 10 yrs

Carbon Dioxide (CO <sub>2</sub> )	20% volume
SENSOR	
Туре	Infrared
Standard Range	0 - 20% volume
Resolution	0.1% volume
Accuracy	$\pm$ 1% volume at STP( 20°C, 101.325 kPa) & time of calibration
Long Term Drift	$\pm$ 1% volume / month @ 20°C (68°F) ambient
Response Time	$t_{90} = > 30 \text{ seconds } @ 20^{\circ}\text{C } (68^{\circ}\text{F})$
Cross Sensitivities	None

# INSTRUMENT

Displayed Resolution	0.1% volume
Warm Up Time @ Switch On	Approx 2 - 2.5 min (warm up delay time)
Recommended Calibration Frequency	2 yrs, 1 yr for best accuracy

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	5 - 95% non-condensing
Operating Life (Estimated)	5 - 10 yrs

Carbon Dioxide (CO <sub>2</sub> )	100% volume
SENSOR	
Туре	Infrared
Standard Range	0 - 100% volume
Resolution	1% of measuring range for readings above 50% of range,
	0.5% of measuring range for readings below 50% of range
Accuracy	$\pm$ 1% volume at STP (20°C, 101.325 kPa) & time of calibration
	$\pm$ 10% volume across temperature and pressure when
	calibrated at altitude
Long Term Zero Drift	± 1% volume / month @ 20°C (68°F) ambient
Response Time	$t_{90} = < 30$ seconds @ 20°C (68°F) ambient
Cross Sensitivities	None

Displayed Resolution	0.1% volume
Warm Up Time @ Switch On	Approx 3 min (warm up delay time)
Recommended Calibration Frequency	6 months

# **ENVIRONMENTAL**

Operating Temperature	-20°C to 50°C (-4°F to 122°F)
Operating Humidity	0 - 95% non-condensing
Operating Life (Estimated)	> 5 yrs

Carbon Monoxide (CO)	50 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 50 ppm
Resolution	0.5 ppm
Accuracy	No data available
	zero: 0.2 ppm equivalent change / yr in clean air.
Long Term Drift	Sensitivity: 3% change / yr in clean air (value based on twice
	per month test)
Response Time	t <sub>90</sub> =< 25 seconds fr 0 - 400 ppm
	$H_2 S @ 20 \text{ ppm} = < 0.1 \text{ ppm}$
Cross Sensitivities	$NO_2 @ 10 \text{ ppm} = < 0.1 \text{ ppm}$
Closs Selisitivities	$Cl_2 @ 10 \text{ ppm} = < 0.1 \text{ ppm}$
	NO @ 50 ppm = < 5 ppm
	$SO_2 @ 20 \text{ ppm} = < 0.1 \text{ ppm}$
Cross Sensitivities continued	$H_2 @ 20^{\circ}C (68^{\circ}F) @ 400 \text{ ppm} = < 60 \text{ ppm}$
Closs Selisitivities Continued	$C_2H_4 @ 400 \text{ ppm} = < 25 \text{ ppm}$
	NH <sub>3</sub> @ 20 ppm = < .01 ppm

# INSTRUMENT

Displayed Resolution	1 ppm
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	2 yrs, 1 yr for best accuracy

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	15 - 90% non-condensing
Operating Life (Estimated)	2 - 3 yrs

Carbon Monoxide (CO)	50 ppm
SENSOR	
Туре	Electrochemical (for use in H <sub>2</sub> background environment)
Standard Range	0 - 50 ppm
Resolution	0.5 ppm
Accuracy	No data available
	zero: 0.2 ppm equivalent change / yr in clean air.
Long Term Drift	Sensitivity: 3% change / yr in clean air (value based on twice
	per month test)
Response Time	$t_{90} = < 30$ seconds
	$H_2$ @ 900 ppm in 900 ppm CO @ 10°C (50°F) = < 2 ppm
	$H_{2}^{-}$ @ 900 ppm in 900 ppm CO @ 20°C (68°F) = < 4 ppm
	$H_{2}^{-}$ @ 900 ppm in 900 ppm CO @ 30°C (86°F) = < 6 ppm
	$NO_{2}$ @ 10 ppm = < 0.1 ppm
Cross Sensitivities	$Cl_{2}$ @ 10 ppm = < 0.1 ppm
	$N\bar{0} @ 50 \text{ ppm} = < 0.1 \text{ ppm}$
	$SO_{2} @ 20 \text{ ppm} = < 0.1 \text{ ppm}$
	$C_{3}H_{4} @ 400 \text{ ppm} = < 30 \text{ ppm}$
	NH <sub>3</sub> @ 20 ppm = < 0.1 ppm

INSTITUTENT	
Displayed Resolution	0.5 ppm
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	15 - 90% non-condensing
Operating Life (Estimated)	2 - 3 yrs

Chlorine (Cl <sub>2</sub> )	5 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 5 ppm
Resolution	0.02 ppm
Accuracy	No data available
Long Term Drift	zero: < 0.2 ppm equivalent change / yr in clean air with monthly test.  Sensitivity: < 0.4 ppm change / month in clean air with twice monthly test

Response Time	$t_{so} = $ < 40 seconds fr 0 - 5ppm (diffusion)
	H <sub>2</sub> S @ 20 ppm= < -40 ppm
	NO <sub>3</sub> @ 10ppm= 100 ppm
	NO @ 50 ppm = < 0.5 ppm
Cross Sensitivities	SO <sub>2</sub> @ 20 ppm = < -2.5 ppm
	CO @ 400 ppm = < 0.1 ppm
	H <sub>2</sub> @ 400 ppm = < 0.1 ppm
	$C_{5}H_{4} @ 400 \text{ ppm} = < 0.1 \text{ ppm}$

Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

# **ENVIRONMENTAL**

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	15 - 90% non-condensing
Operating Life (Estimated)	2 - 2.5 yrs

Chlorine Dioxide (ClO <sub>2</sub> )	1 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 1 ppm
Resolution	0.02 ppm
Accuracy	No data available
Long Term Drift	< 5% / 6 months
Response Time	$t_{50} = $ < 20 seconds calc fr 2 minute exposure time
nesponse fille	$t_{90} = < 120$ seconds calc fr 2 minute exposure
	Alcohols @ 1,000 ppm = 0 ppm
	CO @ 100 ppm = 0 ppm
Cross Sensitivities	Cl <sub>2</sub> @ 1ppm = 0.6 ppm
	$0_{3}^{2}$ @ 0.25 ppm = 0.7 ppm
	$H_{3} @ 3,000 \text{ ppm} = 0 \text{ ppm}$
	$H_3^2$ S @ 20 ppm = -5 ppm

#### **INSTRUMENT**

Displayed Resolution	0.01 ppm (10 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

# **ENVIRONMENTAL**

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	15 - 95% non- condensing
Operating Life (Estimated)	2 yrs

Combustibles	100% LEL
SENSOR	
Туре	Catalytic Pellistor
Standard Range	0 - 100% LEL
Resolution	1% LEL
Accuracy	No data available
Long Term Drift	Minimal
Response Time	$t_{so} = < 10$ seconds
Cross Sensitivities	Responds to most flammable gases & vapours

# INSTRUMENT

Displayed Resolution	1% LEL
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	1 yr for best performance

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	10 - 90% non- condensing
Operating Life (Estimated)	5 yrs +

Combustibles	5% volume CH <sub>4</sub>
SENSOR	·
Туре	Infrared
Standard Range	0 - 5% vol CH <sub>4</sub>
Resolution	0.1% vol CH <sub>4</sub>
Accuracy	No data available
Long Term Drift	$\pm$ 1% FSD / mth @ 20°C (68°F) ambient, (max $\pm$ 3% of full scale / yr)
Response Time	$t_{oo} = <30$ seconds @ 20°C (68°F) ambient
Cross Sensitivities	None

Displayed Resolution	0.1% volume
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	1 year

#### **ENVIRONMENTAL**

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	5 – 95% non-condencing
Operating Life (Estimated)	5 - 10 yrs

Ethylene (C <sub>2</sub> H <sub>4</sub> )	200 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 200 ppm
Resolution	1 ppm
Accuracy	No data available
Long Term Drift	< 5% / month
Response Time	$t_{so} = < 100$ seconds
Cross Sensitivities	co = < 60%

# INSTRUMENT

Displayed Resolution	1 ppm
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	15 - 90% non- condensing
Operating Life (Estimated)	2 - 3 yrs

Ethylene Oxide (C <sub>2</sub> H <sub>4</sub> O)	20 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 20 ppm
Resolution	0.1 ppm
Accuracy	No data available
Long Term Drift	< 5% signal loss / yr
Response Time	$t_{90} = < 120$ seconds

Cross Sensitivities Ethanol $\approx 55\%$ Toluene $\approx 20\%$ Methyl-ethyl-ketone $\approx 10\%$ C0 $\approx 40\%$
--

Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	15 - 90% non- condensing
Operating Life (Estimated)	2 - 3 yrs

Fluorine (F <sub>2</sub> )	2 ppm				
SENSOR					
Туре	Electrochemical				
Standard Range	0 - 2 ppm				
Resolution	< 0.02 ppm @ 20°C (68°F)				
Accuracy	No data available				
Long Term Drift	< 5% / month				
Response Time	$t_{90} = < 80$ sec calc fr 4 minute exposure with 1 ppm Cl <sub>2</sub>				
	Alcohols @ 1,000 ppm = 0 ppm				
	$AsH_{3} @ 0.2 \text{ ppm} = -0.03 \text{ ppm}$				
	Br = yes; n/d				
	$CO_{2}$ @ 5,000 ppm = 0 ppm				
	CO @ 100 ppm = 0 ppm				
	Cl <sub>2</sub> @ 1 ppm = 1.4 ppm				
	$B_{3}H_{6} @ 0.25 \text{ ppm} = -0.01 \text{ ppm}$				
	Hydrocarbons @ % range = 0 ppm				
Cross Sensitivities	HCl @ 5 ppm = -7 ppm				
	$H_{2} @ 10,000 \text{ ppm} = 0 \text{ ppm}$				
	HCN @ 1  ppm = -0.05  ppm				
	$H_3S @ 1 ppm = -2 ppm$				
	$N_{2}^{-}$ @ 100% = 0 ppm				
	$NO_{,}$ @ 10 ppm = 8 ppm				
	$0_{3} @ 0.25 \text{ ppm} = 0.3 \text{ ppm}$				
	$PH_3 @ 0.3 \text{ ppm} = \text{approximately -0.1 ppm; n/d}$				
	SO <sub>2</sub> @ 20 ppm = -0.2 ppm				

	-				NIT
IN	31	ΚI	JΛ	Λŀ	NT

Displayed Resolution	0.01 ppm (10 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

#### **ENVIRONMENTAL**

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	15 - 90% non- condensing
Operating Life (Estimated)	1.5 - 2 yrs

Formaldehyde (CH <sub>2</sub> 0)	5 ppm	See Important Notes # 4
SENSOR		
Туре	Electrochemical	
Standard Range	0 - 5 ppm	
Resolution	0.01 ppm	
Accuracy	No data available	
Long Term Drift	< 2% signal loss / month	
Response Time	$t_{so} = < 80 \text{ sec}$	
	$H_{2} = 1 - 3\%$	
Cross Sensitivities	$\tilde{C0} = 10 - 18\%$	
	Interference from other redu	cing gases such as alcohol.

#### INSTRUMENT

Displayed Resolution	0.01ppm (10 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

Operating Temperature	0°C to 50°C (32°F to 122°F)	
Operating Humidity	15 - 90% non- condensing	
Operating Life (Estimated)	2 - 3 yrs in air	

1,000 ppm
Electrochemical
0 - 1,000 ppm
2 ppm
No data available
< 2% / months
$t_{q_0} = < 90$ seconds
$CO @ 300 \text{ ppm} = \le 60 \text{ ppm}$
$H_2 S @ 15 \text{ ppm} = < 3 \text{ ppm}$
$SO_2 @ 5 \text{ ppm} = 0 \text{ ppm}$
NO @ 35 ppm ≈ 10 ppm
$NO_{2}$ @ 5 ppm = 0 ppm
$Cl_{2}$ @ 1 ppm = 0 ppm
HCN @ 10 ppm ≈ 3 ppm
HCl @ 5 ppm = 0 ppm
$C_2H_4$ @ 100 ppm $\approx$ 80 ppm

Displayed Resolution	1 ppm
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	15 - 90% non- condensing
Operating Life (Estimated)	2 yrs +

Hydrogen Chloride (HCl)	30 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 30 ppm
Resolution	< 0.7 ppm @ 20°C (68°F)
Accuracy	No data available
Long Term Drift	< 3% / month
Response Time	$t_{90} = < 70$ seconds calc fr 4 minute exposure
	Alcohols @ 1,000 ppm = 0 ppm
Cross Sensitivities	$NH_3 @ 100 \text{ ppm} = 0 \text{ ppm}$
	$AsH_{3} @ 0.2 \text{ ppm} = 0.7 \text{ ppm}$
	$CO_{2} @ 5,000 \text{ ppm} = 0 \text{ ppm}$

	CO @ 100 ppm = 0 ppm	
	$Cl_{3}$ @ 5 ppm = $< \pm 0.1$ ppm	
	Hydrocarbons @ % range = 0 ppm	
	$H_{2}$ @ 10,000 ppm = 0 ppm	
	HCN @ 20 ppm = 7 ppm	
Cross Sensitivities continued	$H_{2}S @ 20 \text{ ppm} = 60 \text{ ppm}$	
	NO @ 100 ppm = 45 ppm	
	$N_{2} @ 100\% = 0 \text{ ppm}$	
	$NO_{3}$ @ 10 ppm = $< \pm 0.5$ ppm	
	$PH_{3}^{2} @ 0.1 = 0.3 \text{ ppm}$	
	SO <sub>3</sub> @ 20 ppm = 8 ppm	

Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	15 - 95% non- condensing
Operating Life (Estimated)	2 yrs

Hydrogen Cyanide (HCN)	30 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 30 ppm
Resolution	0.2 ppm
Accuracy	No data available
Long Term Drift	< 5% / month
Response Time	$t_{90} = < 50$ seconds calc fr 2 minute exposure
	Alcohols @ 1,000 ppm = 0 ppm
	$CO_{5} @ 5,000 \text{ ppm} = 0 \text{ ppm}$
	CO @ 100  ppm = 0  ppm
	Hydrocarbons @ % range = 0 ppm
Cross Sensitivities	$H_{2} @ 10,000 \text{ ppm} = 0 \text{ ppm}$
Cross Sensitivities	NO @ 100  ppm = -5  ppm
	$NO_{3}$ @ 10 ppm = -7 ppm
	$H_{3}S @ 20 \text{ ppm} = 0 \text{ ppm}$
	(short gas exposure in minute range;
	after filter saturation: ca. 40 ppm reading)

Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

#### **ENVIRONMENTAL**

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	15 - 95% non- condensing
Operating Life (Estimated)	1.5 - 2 yrs

Hydrogen Fluoride (HF)	10 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 10 ppm
Resolution	0.2 ppm @ 20°C (68°F)
Accuracy	No data available
Long Term Drift	< 10% in 6 months
Response Time	$t_{90} = < 90$ seconds calc fr 4 minute exposure
	$C_1H_4O_2$ @ 100 ppm = 100 ppm
	Alcohols @ 1,000 ppm = 0 ppm
	$CO_2 @ 5,000 \text{ ppm} = 0 \text{ ppm}$
	CO @ 100 ppm = 0 ppm
Cross Sensitivities	$Cl_2@1 \text{ ppm} = 0.7 \text{ ppm}$
	Hydrocarbons @ % range = 0 ppm
	$H_{2} @ 3,000 \text{ ppm} = < 1 \text{ ppm}$
	HCI @ 10 ppm = 6 ppm
	SO <sub>3</sub> @ 20 ppm = 16 ppm

#### INSTRUMENT

Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	15 - 90% non- condensing
Operating Life (Estimated)	1.5 yrs +

Hydrogen Sulphide (H <sub>2</sub> S)	50 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 50 ppm
Resolution	< 0.05 ppm
Accuracy	No data available
Long Term Drift	Zero: $<$ 0.1 ppm equivalent change / yr in clean air
	Sensitivity: < 4% change / yr in clean air with monthly test
Response Time	$t_{90} = < 25$ seconds fr 0 - 20 ppm
	$NO_{2}$ @ 10 ppm = < -20 ppm
	Cl <sub>2</sub> @ 10 ppm = -25 ppm
	NO @ 50  ppm = < 4  ppm
Cross Sensitivities	SO <sub>2</sub> @ 20 ppm = < 10 ppm
Cross Sensitivities	CO @ 400 ppm = < 1.5 ppm
	$H_{3} @ 400 \text{ ppm} = < 0.2 \text{ ppm}$
	$C_{2}H_{A} @ 400 \text{ ppm} = < 0.5 \text{ppm}$
	$NH_{3}$ @ 20 ppm = < 0.1 ppm

Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

ENTINONNENTAL	
Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	15 - 90% non- condensing
Operating Life (Estimated)	2 - 3 vrs

Nitrogen Dioxide (NO <sub>2</sub> )	5 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 5 ppm
Resolution	0.1 ppm
Accuracy	No data available
Long Term Drift	< 2% signal loss / month
Response Time	$t_{90} = < 25$ seconds
	$H_2^{S}$ @ 20 ppm = < -40 ppm
	$Cl_2$ @ 10 ppm = 100 ppm
	NO @ 50  ppm = < 0.5  ppm
Cross Sensitivities	SO <sub>2</sub> @ 20 ppm = < -2.5 ppm

	CO @ 400 ppm = < 0.1 ppm H <sub>2</sub> @ 400 ppm = < 0.1 ppm
Cross Sensitivities continued	$C_2^2 H_4 @ 50 \text{ ppm} = < 0.1 \text{ ppm}$
	$NH_{3}$ @ 20 ppm = < 0.1 ppm
	$CO_{3}$ @ 5% volume = < 0.1 ppm

Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	15 - 90% non- condensing
Operating Life (Estimated)	2 - 3 yrs

Nitric Oxide (NO)	100 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 100 ppm
Resolution	< 0.2 ppm
Accuracy	No data available
Long Term Drift	Zero: 0.5 ppm equivalent change fr $-20^{\circ}$ C to $20^{\circ}$ C ( $-4^{\circ}$ F to $68^{\circ}$ F), $1$ – 3 ppm equivalent change $20^{\circ}$ C to $50^{\circ}$ C ( $68^{\circ}$ F to $122^{\circ}$ F)  Sensitivity: $101$ – $105\%$ output change @ $50$ ppm btw $20^{\circ}$ C ( $68^{\circ}$ F) & $50^{\circ}$ C ( $122^{\circ}$ F)
Response Time	$t_{q_0} = $ < 20 seconds fr 0 - 50 ppm
Cross Sensitivities	$H_2S @ 20 \text{ ppm} = < 30 \text{ ppm}$ $NO_2 @ 50 \text{ ppm} = < 5 \text{ ppm}$ $CI_2 @ 10 \text{ ppm} = < 15 \text{ ppm}$ $SO_2 @ 20 \text{ ppm} = < 3 \text{ ppm}$ $H_2 @ 400 \text{ ppm} = < 0.1 \text{ ppm}$ $CO @ 400 \text{ ppm} = < 0.1 \text{ ppm}$ $NH_3 @ 20 \text{ ppm} = < 0.1 \text{ ppm}$ $CO_2 @ 5\% \text{ volume} = < 0.1 \text{ ppm}$

Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

# **ENVIRONMENTAL**

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	15 - 90% non- condensing
Operating Life (Estimated)	2 - 3 yrs

Oxygen (O <sub>2</sub> )	25% volume
SENSOR	
Туре	Electrochemical
Standard Range	0 - 25% volume
Resolution	0.1% volume
Accuracy	No data available
Long Term Drift	< 1% change in output over 3 months
Response Time	$t_{90} = < 15$ seconds fr 0 - 20.9%
Cross Sensitivities	CO <sub>2</sub> sensitivity: 0.1% change in O <sub>2</sub> reading per % CO <sub>2</sub> in 5% CO <sub>3</sub>

#### INSTRUMENT

Displayed Resolution	0.1% volume
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

Operating Temperature	0°C to 55°C (32°F to 131°F)
Operating Humidity	< 0.7% fr 0 - 95% RH @ 40°C (104°F)
Operating Life (Estimated)	3 yrs

Ozone (0,)	1 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 1 ppm
Resolution	< 0.02 ppm @ 20°C (68°F)
Accuracy	No data available
Long Term Drift	< 10% / 6 months @ 20°C (68°F) and 30 - 50% RH
Response Time	$t_{90} = < 60$ seconds calc from 3 minute exposure @ 30 cc / min flow
	Br, $I_2$ = yes; n/d
	$CO_{2}^{-}$ @ 5,000 ppm = 0 ppm
	CO @ 100 ppm = 0 ppm
Cross Sensitivities	Cl <sub>2</sub> @ 1 ppm = 1.2 ppm
	$N_2H_4$ @ 3 ppm = -3 ppm
	$H_2 @ 3,000 \text{ ppm} = 0 \text{ ppm}$
	$H_2S @ 20 \text{ ppm} = -1.6 \text{ ppm}$
	$N_2 @ 100\% = 0 \text{ ppm}$
	$NO_2$ @ $10$ ppm = 6 ppm
INSTRUMENT	
Displayed Resolution	0.01 ppm (10 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months
ENVIRONMENTAL	
Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	15 - 90% non-condensing

1.5 - 2 yrs

Operating Life (Estimated)

Phosphine (PH <sub>3</sub> )	5 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 5 ppm
Resolution	Lower detection limit < 30 ppb
Accuracy	No data available
Long Term Drift	< 5% / 6 months
Response Time	$t_{q_0} = < 30 \text{ seconds}$
	CO @ 85 ppm = 0 ppm
	$H_{2} @ 3,100 \text{ ppm} = 0 \text{ ppm}$
	$N_{0}^{0}$ @ 10 ppm = 2 ppm
	$C_3H_5$ OH @ 25,000 ppm = 0 ppm
	H,S @ 18 ppm = 13 ppm
	$S_{0}^{2}$ @ 18 ppm = 6.5 ppm
Constantiniai o	$Cl_{3}@0.85 \text{ ppm} = 0.29 \text{ ppm}$
Cross Sensitivities	HCI @ 7.8 ppm = 1.2 ppm
	HF @ 7.2 ppm = 0 ppm
	HCN @ 12.6 ppm = 0.84 ppm
	$SiH_4 @ 4.3 \text{ ppm} = 0.84 \text{ ppm}$
	H,Se @ 0.8 ppm = 0.29 ppm
	$B_{3}^{2}H_{6} @ 0.2 \text{ ppm} = 0.34 \text{ ppm}$
	AsH <sub>3</sub> @ 0.2 ppm = 0.16 ppm
INSTRUMENT	
Displayed Resolution	0.01 ppm (10 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months
ENVIRONMENTAL	
Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	10 - 95%
Operating Humidity	non- condensing
Operating Life (Estimated)	2 yrs

Silane (SiH₄)	20 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 20 ppm
Resolution	0.05 ppm
Accuracy	No data available
Long Term Drift	< 5% / 6 months
Response Time	$t_{on} = < 60$ seconds calc fr 2 min exposure
Cross Sensitivities	C0 @ 85 ppm = 0 ppm  H <sub>2</sub> @ 3,100 ppm = 0 ppm  NO <sub>2</sub> @ 10 ppm = 2.3 ppm  C <sub>3</sub> H <sub>5</sub> OH @ 25,000 ppm = 0 ppm  H <sub>2</sub> S @ 18 ppm = 8 ppm  SO <sub>2</sub> @ 18 ppm = 7.4 ppm  Cl <sub>2</sub> @ 0.85 ppm = 0.1 ppm  HCl @ 8 ppm = 0.45 ppm  HF @ 7.2 ppm = 0 ppm  HCN @ 12 ppm = 0.77 ppm  AsH <sub>3</sub> @ 0.16 ppm = 0.2 ppm  H <sub>2</sub> Se @ 0.8 ppm = 0.2 ppm  B <sub>2</sub> H <sub>6</sub> @ 0.2 ppm = 0.27 ppm  PH <sub>3</sub> @ 0.2 ppm = 0.35 ppm

Displayed Resolution	0.01 ppm (10 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

Operating Temperature	0°C to 40°C (32°F to 104°F)	
Operating Humidity	20 - 95% non- condensing	
Operating Life (Estimated)	1.5 yrs	

Sulphur Dioxide (SO <sub>2</sub> )	20 ppm	
SENSOR		
Туре	Electrochemical	
Standard Range	0 - 20 ppm	
Resolution	< 0.1 ppm	
Accuracy	No data available	
Long Term Drift	< 2% change / month in clean air	
Response Time	$t_{q_0} = < 25$ seconds fr 0 - 10 ppm	
	$H_3S @ 20 \text{ ppm} = < 0.1$	
	$N_{0}^{-}$ @ 10 ppm = < -130 ppm	
	Cl <sub>2</sub> @ 10 ppm = < -40 ppm	
Cross Sensitivities	$NO @ 50 \text{ ppm} = < \pm 2 \text{ ppm}$	
cross sensitivities	CO @ 400 ppm = < 1.6 ppm	
	$H_{2}$ @ 400 ppm = < 0.3 ppm	
	$C_{2}H_{4} @ 400 \text{ ppm} = < 40 \text{ ppm}$	
	$NH_{3}$ @ 20 ppm = < 0.1 ppm	
INSTRUMENT		
Displayed Resolution	0.1 ppm (100 ppb)	
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)	
Recommended Calibration Frequency	6 months	
ENVIRONMENTAL		
Operating Temperature	0°C to 40°C (32°F to 104°F)	
Operating Humidity	15 - 90% non- condensing	
Operating Life (Estimated)	2 yrs	

Total Volatile Organic Compo	ound (TVOC) 30 ppm	
SENSOR		
Туре	Photolonization Detector	
Standard Range	0 - 30 ppm	
Resolution	0.02 ppm	
Accuracy	No data available	
Long Term Drift	< 2% change / month in clean air	
Response Time	$t_{90} = < 3$ seconds	
Cross Sensitivities	Many chemicals & gases. Refer to manual.	
INSTRUMENT		
Displayed Resolution	0.01 ppm (10 ppb)	
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)	
Recommended Calibration Frequency	monthly to 4 months (usage dependent)	
ENVIRONMENTAL		
Operating Temperature	0°C to 40°C (32°F to 104°F)	
Operating Humidity	5 - 95% non- condensing	
Operating Life (Estimated)	5 yrs (excluding replacable lamp & electrode stack)	

Total Volatile Organic Co	mpound (TVOC)	300 ppm
SENSOR		
Туре	Photolonization Detect	or
Standard Range	0 - 300 ppm	
Resolution	0.1 ppm	
	(100 ppb)	
Accuracy	No data available	
Long Term Drift	< 2% change / month	in clean air
Response Time	$t_{q_0} = <3$ seconds	
Cross Sensitivities	Many chemicals & gase	es. Refer to manual.
INSTRUMENT		
Displayed Resolution	1 ppm	
Warm Up Time @ Switch On	Approximately 2 - 2.5 r	nin (instrument warm up delay time)

LIVINONNENTAL	
Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	5 - 95% non- condensing
Operating Life (Estimated)	5 yrs (excluding replacable lamp & electrode stack)
operating the (Estimateu)	J yl3 (chelduling replacable lamp & electrode stack)

monthly to 4 months (usage dependent)

Warm Up Time @ Switch On **Recommended Calibration Frequency** 

ENVIDANMENTAL

Denti mlete Comon	DU2 F DU10	
Particulate Sensor	PM2.5 or PM10	
SENSOR		
Туре	Optical - laser LED and photosensor	
Particle Size Range	0.3 - 10 μm	
Resolution	1 μg/m³	
Detection Range	1 - 999 μg/m³	
Detection Error		
< 100 μg/m3		
> 100 μg/m3	±15 %	
Response Time	10 sec	
	Build up of dust adhered to the vents or the inside of the	
	sensor should be cleaned with a vacuum or compressed air	
	duster.	
	The density of the particles being monitored is relevant to the	
Sensitivities	accuracy of the sensor.	
Sensitivities	Sensor may be affected by noise generating equipment such	
	as an electric dust collector or power supply line.	
	The sensor may be affected by vibration or mechanical	
	oscillation.	
	Avoid adhesive particles such as oil.	
INSTRUMENT		
Displayed Measurement	PM2.5 or PM10	
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)	
Recommended Calibration Frequency	n/a	
ENVIRONMENTAL		
Operating Temperature	-10°C to 50°C (14°F to 122°F)	
Operating Humidity	35 - 85% non- condensing	
	5+ years	
Operating Life (Estimated)	(depends on application and atmosphere, continuous mode	
	or intermittent use)	

#### Additional chemical symbols not defined previously:

Br Bromine C<sub>2</sub>H<sub>2</sub> Acetylene B<sub>2</sub>H<sub>6</sub> Diborane

CIF<sub>3</sub> Chlorine Trifluoride C<sub>3</sub>H<sub>8</sub>O Isopropyl Alcohol

#### **Important Notes:**

- Some sensors may be calibrated with correlation gases. If you prefer to have specific sensors
  calibrated with the target gas, contact our factory for availability and extra costs. Customer will have
  to bear the cost of the full cylinder of specialty gas plus incoming dangerous goods freight and take
  ownership of the cylinder of gas remaining.
- 2. These specifications have been developed from data considered accurate at the time. No warranty is implied or suggested based on this data. We accept no responsibility for errors or omissions.
- 3. Critical Environment Technologies Canada Inc. reserves the right to make design and specification changes without prior notice.
- 4. Formaldehyde sensor has high cross sensitivity to Carbon Monoxide, Alcohol & Hydrogen.
- 5. Combustible (flammable) gas sensors (catalytic) can be calibrated for a number of target gases. Please specify the target gas desired & we will evaluate your request.
- 6. Chlorine, Fluorine, Hydrogen Chloride, Hydrogen Fluoride and Ozone gas sensors are not available with the YESAIR Pump model.

# Critical Environment Technologies Canada Inc.

Unit 145, 7391 Vantage Way, Delta, BC, V4G 1M3, Canada

Toll Free: +1.877.940.8741 Tel: +1.604.940.8741

Fax: +1.604.940.8745

www.critical-environment.com