



CERTIFICATE OF CALIBRATION COMPLIANCE

This is to certify that the CPS scales models listed have been individually calibrated to respond to the published minimum accuracy levels using one of the following NIST traceable standards:

Class F 50 lb weight **S/N's TP3060, 7816, OJOH**
Class F 100 lb weight **S/N's 2390, 1QNZ**

CPS scales covered by this certificate:

CCD110
CC50, CC110
CC200
CC220, CC220E
CC800A
CC220W, CC240W, CC840W

CC240RF, CC840RF
CC100, CC100E
CC700 series
CC800 series
CC1000

A handwritten signature in black ink, appearing to read 'Richard Holby', written over a horizontal line.

Vice President of Engineering
CPS Products, Inc.

August 24, 2017

Date.

CPS WORLD HEADQUARTERS

1010 East 31st Street,
Hialeah, Florida 33013, USA
Tel: 305-687-4121, 1-800-277-3808
Fax: 305-687-3743
E-mail: info@cpsproducts.com
Website: www.cpsproducts.com